Avoiding acidic region streaking in two-dimensional gel electrophoresis: Case study with two bacterial whole cell protein extracts

ARNAB ROY¹, UMESH VARSHNEY² and DEBNATH PAL^{1,*}

¹Bioinformatics Centre, Supercomputer Education and Research Centre and ²Department of Microbiology and Cell Biology, Indian Institute of Science, Bangalore 560 012, India

*Corresponding author (Fax, +918023602648; Email, dpal@serc.iisc.in)

Acidic region streaking (ARS) is one of the lacunae in two-dimensional gel electrophoresis (2DE) of bacterial proteome. This streaking is primarily caused by nucleic acid (NuA) contamination and poses major problem in the downstream processes like image analysis and protein identification. Although cleanup and nuclease digestion are practiced as remedial options, these strategies may incur loss in protein recovery and perform incomplete removal of NuA. As a result, ARS has remained a common observation across publications, including the recent ones. In this work, we demonstrate how ultrasound wave can be used to shear NuA in plain ice-cooled water, facilitating the elimination of ARS in the 2DE gels without the need for any additional sample cleanup tasks. In combination with a suitable buffer recipe, IEF program and frequent paper-wick changing approach, we are able to reproducibly demonstrate the production of clean 2DE gels with improved protein recovery and negligible or no ARS. We illustrate our procedure using whole cell protein extracts from two diverse organisms, *Escherichia coli* and *Mycobacterium smegmatis*. Our designed protocols are straightforward and expected to provide good 2DE gels without ARS, with comparable times and significantly lower cost.

[Roy A Varshney U and Pal D 2014 Avoiding acidic region streaking in two-dimensional gel electrophoresis: Case study with two bacterial whole cell protein extracts. *J. Biosci.* **39** 631–642] DOI 10.1007/s12038-014-9453-9

1. Introduction

Since O'Farrell's paper in 1975 (O'Farrell 1975), two-dimensional gel electrophoresis (2DE) has become a useful tool in the analysis of complex protein mixtures, especially after the popularization of mass-spectrometry-based methods to identify proteins (Henzel *et al.* 1993; James *et al.* 1993; Pappin *et al.* 1993; Rabilloud *et al.* 2010; Yates *et al.* 1993). Its capacity to resolve thousands of proteins in a single gel and to detect post- and co-

translational modifications (Lopez 2007; Rabilloud 2002) has made 2DE a popular instrument in modern proteomics (Butt *et al.* 2012; Hong *et al.* 2013; Pal *et al.* 2013; Rajani *et al.* 2012; Sun *et al.* 2013; Thiede *et al.* 2013; Xiao *et al.* 2012; Zhang *et al.* 2013) as well as in metabolomics research (Chen *et al.* 2012; Kamthan *et al.* 2012; Sun *et al.* 2012; Zhang *et al.* 2012). Moreover, contemporary 2DE instruments' ability to run dozens of 2DE experiments in parallel offers improved reliability and reproducibility (Rabilloud *et al.* 2010).

Keywords. Paper wick; sample preparation; sonication; streaking; two-dimensional electrophoresis

Abbreviations used: ARS, acidic region streaking; dd-Milli-Q, double distilled Milli-Q water; MC-tube, microcentrifuge tube; NuA, nucleic acid; RB, rehydration buffer; WCPE, whole cell protein extract

Supplementary materials pertaining to this article are available on the *Journal of Biosciences* Website at <http://www.ias.ac.in/jbiosci/sep2014/supp/Roy.pdf>

Ideally proteins should appear on a 2DE gel as localized concentric spots. However, artifacts are common problems in 2DE. One such very common artifact is streaking (Cremer and Van de Walle 1985; Hatairaktham *et al.* 2013; Havanapan and Thongboonkerd 2009; Jiang *et al.* 2011; O'Farrell 1975; Qi *et al.* 2013). A streaking on a 2DE gel may be defined as a horizontal or a vertical stretch of stained region with intensity greater than the local background and can conceal a spot. Streaking poses great obstacles during image analysis and confers ambiguity during protein identification. The problem of horizontal streaking at basic region and vertical streaking is more or less answered (Bai *et al.* 2005; Görg *et al.* 1987,1997, 2000; Havanapan and Thongboonkerd 2009; O'Farrell 1975; Olsson *et al.* 2002; Pennington *et al.* 2004). But streaking at the acidic end as such is relatively unattended (Antonioli *et al.* 2009; Herbert *et al.* 2006), where there is scope for major improvement. Acidic region streaking (ARS) generally starts from about pH 5 extending towards lower pH region smearing especially for high molecular weight proteins. Given that a large number of proteins (more than 17% of the *E. coli* proteins) have pI below pH 5, ARS may obscure lot of important information. On overview of dramatically varying quality of 2D gels in terms of streaking and number of clear spots can be seen in archived gel images at EcoProDB(Yun *et al.* 2007).

Among the many factors that contribute to streaking, nucleic acid (NuA) contamination is among the most frequent cause for ARS (Antonioli *et al.* 2009; Rabilloud 1996; Westermeier and Naven 2002). Although treatment with nucleases and/or selective protein precipitation-resolubilization (cleanup) are suggested to eliminate NuA contamination (Antonioli *et al.* 2009; Rabilloud 1996), these are prone to errors due to attenuated nuclease activity in strong denaturing buffer, inadequate resolubilization of proteins, or risk of leftover contamination from enzymes and acids. Not surprisingly, ARS is commonly seen in contemporary 2DE gels, even with the simplest 2DE sample, such as the *E. coli* whole cell protein extract (WCPE) (Liao *et al.* 2011; Nakata *et al.* 2010; Qi *et al.* 2013; Ramachandran *et al.* 2012; Riley *et al.* 2012). Efforts have been made on improving protein recovery without eliminating streaking (Valente *et al.* 2012), which may result in loss of potentially important expression information. Furthermore, any additional step adds to the already multi-step protocol of 2DE, potentially affecting protein recovery and the protein profile itself (Rabilloud 1996). However, for some samples (e.g., plant or plasma samples), cleanup and/or multiple depletion cycles becomes essential in order to get any meaningful profile from 2DE gels (Damerval *et al.* 1986). On the other hand, *E. coli* and prokaryotic WCPE in general have low salt content, giving us an opportunity to simplify the sample preparation step with the aim to maximize protein recovery without disturbing the protein profile.

In this article, a simple solution for ARS is reproducibly demonstrated on *E. coli* and *M. smegmatis* 2DE gels, for which no exclusive sample cleanup was performed. The whole cell lysate was prepared by sonication of the cells in pure water (rather than in urea/thiourea buffer), under fixed temperature of an ice-water bath. A suitable buffer, IEF program, and a modified paper wick usage, yielded improved 2DE gels with negligible or no streaking. An indirect protein estimation appropriate for 2DE gel samples was used, as an alternative for costly 2DE-quantification kits. Our designed protocols are expected to provide 2DE gels without ARS, at less time and lower cost, with improved protein recovery and quantitative protein profile.

2. Materials and methods

WCPE samples from *E. coli* and *Mycobacterium smegmatis* were prepared using known protocols, and our modified protocol (figure 1; supplementary figure 1). Parameters of 2DE experiment were optimized iteratively. Final protocols were repeated at least thrice for reliability. We compared our results against few recently published *E. coli* and *M. smegmatis* WCPE 2DE gels.

2.1 Chemicals and reagents

Double distilled Milli-Q water (dd-Milli-Q) was used where high quality water was preferable. Urea, DTT, BSA, Tris, SDS, acrylamide, bis-acrylamide, EDTA, β -Mercaptoethanol, Tween-80, RNase and bromophenol blue were purchased from Sigma-Aldrich, St Louis, Mo, USA. Thiourea was purchased from S. D. Fine-Chem Limited, (Mumbai). CHAPS was purchased from G-Biosciences, St Louis, Mo, USA. IPG strips, mineral oil and Bio-Lyte 3/10 ampholyte were purchased from Bio-Rad, Hercules, CA, USA. Sample clean up kit was purchased from Amersham Biosciences Corp., Piscataway, USA and coomassie brilliant blue R-250 was purchased from Amresco Inc., Solon, Ohio. DNase was purchased from SRL Pvt. Ltd. (Maharashtra, India).

2.2 Harvesting cells (supplementary figure 1, steps 1 to 4)

E. coli K-12, *thi1*, *relA1*, *spoT1* (Low 1968) cells grown in Lysogeny broth (10 g/L tryptone, 5 g/L yeast extract (Becton Dickinson India Pvt. Ltd., India) and 10 g/L NaCl) media were collected by centrifugation at 6000g for 10 min. Cells were washed with ice-cold dd-Milli-Q thrice. Washed pellet was stored at -20°C or processed immediately. *Mycobacterium smegmatis* cells were grown in Lysogeny broth with 0.2% Tween-80 and harvested as mentioned above.

Figure 1. A flowchart showing the steps of standard protocol and our protocol for the complete 2DE experiment starting from bacterial cell culture to gel scanning. Following symbols indicate: □ → Steps in protocol; **bold** → steps unique to our protocol; *italics* → steps unique to standard protocol; annotations depicting pros and cons corresponding to specific steps in the protocol are numbered 1 to 5.

Two methods were employed to prepare the samples; our protocol (OP) and standard protocol (SP). Both are described below.

2.3 Cell disruption (supplementary figure 1, step 5)

OP: For *E. coli* WCPE, cell pellet from 20 mL culture was resuspended in total volume of 500 μ L ice-cold dd-Milli-Q water in a micro-centrifuge tube (MC-tube). Five rounds of sonication were performed with a micro-probe sonicator (Vibra-Cell, Sonics & Materials Inc., Danbury, CT USA), keeping the MC-tube always inside ice. Each round was 40 s long consisting of 1 s on and 1 s off; 1 min cooling was allowed inside ice between two rounds. Intensity of sonication was carefully controlled to prevent frothing. For *M. smegmatis* WCPE, cell pellet from 40 mL culture equivalent was taken and the same procedure was followed as mentioned above, except that, due to its tough cell wall, 12 rounds of sonication were performed.

SP: The rehydration buffer (RB) recipe was 7 M urea, 2 M thiourea, 4% CHAPS, 65 mM DTT and 2% ampholytes. *E. coli* cell pellet from 8 mL culture was resuspended in 500 μ L RB and sonication was done as above.

2.4 Protein extraction (supplementary figure 1, steps 6 to 9)

OP: 200 μ L lysed cell suspension was added to the solid constituents of RB (210 μ g for 7 M urea, 76 μ g for 2 M thiourea and 20 μ g for 4% CHAPS) and mixed at room temperature. When no solid particles are visible, 16.25 μ L of 2 M DTT solution and 10 μ L ampholyte was added. After complete dissolution volume was made up to 500 μ L with dd-Milli-Q water. After removing insoluble cell debris by spinning the suspension at 18,000g for 15 minutes at 18°C, the sample, '2DE extract' was stored at -20°C in small aliquots (100 μ L) or immediately continued to next step. In a separate MC-tube 200 μ L of the lysed suspension was diluted to 500 μ L with ice cold TME buffer (25 mM Tris-HCl pH 8.0, 2 mM β -Mercaptoethanol and 1 mM Na₂EDTA). After spinning at 18,000g for 15 min at 4°C supernatant was collected as 'TME extract'. In OP 500 μ L was taken for sonication, and then 200 μ L was added to buffer recipe. That means actually 2/5th of the 20 mL culture was added to buffer recipe. 2/5th of 20 mL is $(20 \times 2)/5 = 8$ mL.

SP: Lysed cell suspension in RB was centrifuged 18,000g for 15 min at 18°C and supernatant was collected as 2DE extract as mentioned above.

Sample cleanup: Extracted protein sample in RB was cleaned up using either clean up kit or using phenol/chloroform/isoamyl alcohol method. Cleanup kit was used exactly

as mentioned in the instruction manual (Amersham-Biosciences 2003). Briefly, protein was precipitated with the addition of precipitant and co-precipitant. After centrifugation supernatant was discarded. The pellet was washed first with co-precipitant and then with dd-Milli-Q and thereafter with chilled-wash buffer. The pellet was allowed to dry and was subsequently resuspended in RB. Cleanup with phenol/chloroform/isoamyl alcohol was performed as described in the literature (Antonioli *et al.* 2009). In short, the sample in alkaline buffer (50 mM Tris, pH 9.5) was thoroughly mixed with two volumes of phenol/chloroform/isoamyl alcohol (25:24:1) solution and allowed to sit for 5 minutes. After centrifuging at 21,000g at 4°C for 10 min the upper aqueous layer and the bottom organic layer were carefully removed. Excess of acetone was added to wash the protein precipitate. After centrifugation, the pellet was either solubilized in rehydration buffer or proceeded to second round of phenol/chloroform/isoamyl alcohol treatment.

Enzyme digestion: DNase/RNase was added to the sonicated suspension prior to centrifugation. The enzyme solutions contained required amount of ions MgCl₂ for their activity. The mixtures were incubated on ice for 30 minutes (O'Farrell 1975).

2.5 Protein estimation in 2DE extract

Given that RB components interfere with common colorimetric protein essays (Kao *et al.* 2008), we use a parallel approach wherein we estimate the protein quantity in TME extract using the Bradford assay (Bradford 1976) with the help of a spectrophotometer (Genesys 10UV, Thermo Scientific). Standard curve was prepared with a 100 μ g/mL solution of BSA in dd-Milli-Q. Since both the samples, namely 2DE extract and TME extract, had the same amount of lysed cell suspension and was made up to the same volume (section 2.4), they are expected to contain the equivalent amount of protein, assuming homogeneity and neglecting pipetting, instrument error. SDS-PAGE with equal volume load from both samples confirmed (supplementary figure 2, lanes 1 and 2, and supplementary figure 3) the similar amount of protein content in TME extract and 2DE extract. This parallel method can be safely leveraged to suitably estimate the protein content before running 2DE gels as an alternative to the complicated and costly use of 2DE protein estimation kits.

2.6 Rehydration

OP: To run 2DE with 7 cm IPG strip 50 μ g (200 μ g for 17 cm) of 2DE extract (*E. coli* or *M. smegmatis*) was diluted to 125 μ L (300 μ L for 17 cm) with RB and a trace of Bromophenol Blue, mixed thoroughly and kept at 20°C for

half an hour. This mixture was then centrifuged at 18,000g at 20°C for 15 min to remove any insoluble particles. PROTEAN IEF focusing tray (Bio-Rad) was cleaned with dilute SDS solution, gently brushing with a soft nylon brush. The tray was rinsed and brushed under streaming de-ionized water to remove SDS. This process was repeated till no trace of mineral oil was left. Finally the tray was thoroughly rinsed with Milli-Q water and kept for drying inside an airflow hood. The dry tray was placed inside a PROTEAN IEF cell (Bio-Rad). The IPG strip was rehydrated under active rehydration mode (50 V constant) for 12 h as instructed by the instrument manual (Bio-Rad 2011).

SP: IPG strips were rehydrated as above, but in passive mode, without applying any voltage.

2.7 Isoelectric focusing

After rehydration, the oil on the strip was soaked off with a moistened filter paper (supplementary figure 4, steps 2–4). Two paper wicks, cut from WHATMAN filter paper No. 3, were placed on the electrodes of a fresh clean lane of the IEF tray and 10–15 µL dd-Milli-Q was added to the wicks (supplementary figure 4, step 5). Excess water was blotted off with a filter paper (supplementary figure 4, step 6). Thereafter, the strip was placed in the lane, gel-side down on the moistened paper wicks (supplementary figure 4, step 7). IEF was performed with standardized program depicted in supplementary figure 5. The maximum allowed current was set to 50 µA. Paper wicks were changed (supplementary figure 4, steps 4–7) every hour for at least two (for 7 cm) to three (for 17 cm) times after the start of IEF. After this, paper wick changing was continued until the current dropped below 10 µA per gel. Otherwise, the wicks were changed until the final voltage was reached or maximum twice, whichever was earlier.

2.8 Second dimension

Strip was equilibrated in the equilibration buffer I and II (Bio-Rad 2011) for 15 minutes each at room temperature. After this, the strip was rinsed in SDS-running buffer and loaded on the surface of an SDS-PAGE gel. Melted overlay agarose was laid on the strip. SDS-PAGE was run at constant current at 200 V (for 17 cm, at 10 mA for 30 min and then at 16 mA) till the dye front reached the lower bottom of the gel.

2.9 Staining and scanning of gels

Staining was done with coomassie brilliant blue R250 and destaining with 10% acetic acid and 10% methanol solution. Gels were scanned in a Bio-Rad 2DE gel scanner and saved in ‘.mel’ format.

2.10 Agarose gel electrophoresis

Samples were run on 1% agarose (with a pinch of Ethidium bromide) horizontal gels to check for nucleic acid contamination. Gels were run at constant voltage of 90 volts. After the run, gels were scanned inside an UV illuminated scanner.

2.11 Measurement of ARS

Approximate length of horizontal streaks in the acidic region was measured using Gimp image editing software, from published 2DE gel images available in ‘portable document format (PDF)’. We analyzed 2DE gel images, including those from *E. coli* WCPE published in the last one year, emphasizing papers optimizing 2DE gels from *E. coli* WCPE. We analyzed all published 2DE gel images of *M. smegmatis* WCPE accessible to us. ARS was defined as a horizontal stretch of stained line visible on the gel with intensity more than the local background intensity. The distance between two pH ends of a streak was measured in pixel units (x). The length of the longest streak was assigned as (y) pixel units. Therefore, from the pH range information (e.g., if the IEF pH range was 4–7 then a = 4 and b = 7) an approximate streak-length was calculated in terms of pH units using the equation: $\text{Length of streaking} = \frac{x}{y} \times (b - a)$. Number of ARS visible was also recorded; a number 10 meant that the gel had ≥ 10 ARS. 2DE gel images were analyzed with ‘ImageMaster 2D Platinum 7.0’ software; number of spots detected was recorded for comparison.

3. Results

3.1 Effect of ampholyte concentration

For *E. coli* WCPE many variations of buffer recipe have been previously used; however, the standard RB (7 M urea, 2 M thiourea, 65 mM DTT and 4% CHAPS) gave us better results, compared to other compositions (supplementary figure 6). We have checked with ampholyte (pH 3–10) concentrations of 0.5% (figure 2Ai), 1% (data is not shown) and 2% (figure 2Aii). Low ampholyte concentration (0.5%) produced poorly focused gels with horizontal streaking throughout the gel. Increased concentration of ampholyte (2%) showed better focusing and streaking was limited to the acidic region.

3.2 Elimination of NuA contamination

To systematically address the problem of ARS (figure 2Aii) we revisited the issue of non-protein impurity reduction in

WCPE. Standard sample cleanup was performed, which improved focusing, but ARS was visible even with low protein load (supplementary figure 7). It may be noted that relatively high amounts of NuA contamination in *E. coli* WCPE is known to cause ARS (Antonioli *et al.* 2009; Rabilloud 1996; Westermeier and Naven 2002); however, amount of salt, due to its low molar content, is not an issue. Interestingly, digestion with nuclease in RB medium could not completely remove the genomic DNA (compare figure 2B, lanes 2 and 3); neither did it affect the protein recovery much (compare supplementary figure 2, lanes 4 and 5). After cleanup with phenol/chloroform/isoamyl alcohol (Antonioli *et al.* 2009), a substantial amount of NuA was removed (figure 2B, lanes 4 and 5); however, repeated precipitation and resolubilization of WCPE incurred a significant amount of protein loss (compare supplementary figure 2, lanes 4 and 6).

To alleviate the denaturing effect of RB on the activity of nucleases we attempted digestion of NuA in plain water (TME or other buffers were avoided to reduce salt contamination). For that we sonicated cells in water and then added enzymes and the required ingredients. As a result an efficient removal of small-to-medium NuA, as well as the large genomic DNA was observed (figure 2B, lane 6).

In our protocol, we endeavored to prepare the sample without nuclease treatment, which entailed prolonged yet careful sonication in water. Sonication is known to shear large NuA molecules (Westermeier and Naven 2002), while sonication in a denser medium such as RB could not remove NuA. Even increasing number of rounds of sonication in RB did not bring any apparent change in the result, and the large NuA contaminations were still conspicuous in agarose gel (data not shown). Controlled sonication in ice cold water in ice-water-bath followed by addition of RB ingredients and centrifugation resulted in almost similar result as with nuclease treatment (figure 2B, lane 7). Most importantly, the IEF-interfering large genomic DNA was not visible in the agarose gel. Moreover, the protein recovery was maintained similar compared to samples made by standard protocols (compare supplementary figure 2, lanes 1, 2 and 3, and supplementary figure 3).

3.3 IEF optimization and paper-wick changing strategy

We adopted a prolonged IEF program (supplementary figure 5), lengthening the low-voltage initial steps to remove the contaminant ions, as well as the high-voltage final focusing step to compensate for the higher ampholyte concentration (Khoudoli *et al.* 2004). During IEF we replaced the paper wicks that accumulate contaminating ions. Wicks made from WHATMAN filter paper No. 3 improved our results compared to thin paper wicks. Importantly, the excess amount of water on the paper wick when not blotted away as

mentioned above, gives rise to local clusters of horizontal streaks (supplementary figure 8B). Following our sample preparation, we reached the low current (below 10 μ A) very soon, namely, after 2 h for 7 cm strips, and between 3 to 3.5 h for 17 cm strips. So, further changing of paper wicks was not necessary. The procedure gave us clean 2DE gel with more spots with negligible or no ARS (figure 2Ci and supplementary figure 8A). For samples treated with nuclease, some background streaking could be observed in the acidic parts of the gels and spots had smearing at both ends. When nuclease treatment was excluded, gels were highly improved with more focused spots and negligible or no ARS (figure 2Cii) even in the low pH range (figure 2Di). The method is reproducible and assures good protein recovery.

3.4 Test of robustness of the method

To establish the effectiveness of our procedure to eliminate ARS for WCPE from a variety of bacteria, we applied the procedure to *E. coli* and *M. smegmatis*. *E. coli* is from phylum *proteobacteria*, while *M. smegmatis* is from *actinobacteria*. Most importantly, *M. smegmatis* has a hydrophobic cell wall enriched with mycolic acid content. The proteome consists of over 6000 proteins, about 2400 more than that of *E. coli*. Consequently, there are higher number of proteins in the pH 4–7 range; especially clustered around pH 5 (supplementary figure 9A and B). *M. smegmatis* also expresses a higher percentage of hydrophobic proteins than *E. coli* (supplementary figure 9C and D). In spite of the distinct profile of WCPE from the two organisms, our method produced clean 2DE gels, especially for *M. smegmatis* WCPE with large number of spots with negligible or no ARS (figure 2Dii).

3.5 Comparative analysis of ARS

We present a measurement of streaking in terms of approximate pH units as described in section 2.11. We calculated the length of the longest ARS and number of such ARS in gels produced by our protocol (OP) and gels published in few recent articles (figure 3, supplementary figure 10, supplementary tables 1 and 2). It was found that our method produced shortest and lowest number of ARS, if at all. We also compared the number of spots detected in our 2DE gels with the ones which specifically dealt with ARS in 2DE with *E. coli* WCPE.

The work of Antonioli *et al.* discussing elimination of DNA/RNA contamination in *E. coli* WCPE (Antonioli *et al.* 2009), detected an average of 770 spots with Sypro Ruby staining in the pH range 3–10. We were able to detect 666 spots from *E. coli* WCPE on a smaller pH range of pH 4–7 with coomassie R250 staining. It may be noted that Sypro

Figure 2. (A) 2DE images of *E. coli* WCPE on pH 4-7 IPG 17 cm strips: (i) With 0.5% ampholytes and (ii) with 2% ampholytes. (B) Agarose gel electrophoresis images of *E. coli* WCPE. Lane marked (2) correspond to SP, (3) SP followed by DNase/RNase treatment, (4) SP followed by one round of Phenol/chloroform/isoamyl alcohol treatment, (5) SP followed by two rounds of Phenol/chloroform/isoamyl alcohol treatment, (6) OP followed by DNase/RNase treatment, (7) OP alone, (9) SP followed by clean up. Lanes marked 1 and 8 show molecular weight markers. The white band highlighted using an arrow indicates genomic DNA. (C) 2DE images of *E. coli* WCPE on pH 4-7 IPG 17 cm strips: (i) made by OP then DNase/RNase treated and (ii) made by OP with optimized IEF. (D) 2DE image of (i) *E. coli* WCPE on pH3-6 IPG 17 cm strip produced with OP and (ii) *M. smegmatis* WCPE on pH4-7 IPG 17 cm strip produced with our optimized protocol.

Ruby is known to detect twice more spots than coomassie R250 (Chiangjong and Thongboonkerd 2009). In another 2DE optimization work with *E. coli* WCPE, Valente *et al.* detected 626 spots in pH 3–10 with marginal streaking (Valente *et al.* 2012). In case of *M. smegmatis* WCPE, OP yielded 780 2DE spots in pH 4–7 compared to 231 in pH 3–10 (Shires and Steyn 2001) from published 2DE images with no ARS. It may be noted that there are several resources, for example, EcoProDB (Yun *et al.* 2007), that provide comprehensive information on 2D gels and associated proteome, but none have surveyed the streaking present in the archived gels and presented a comparative assessment of quality based on the same.

4. Discussion

4.1 An optimum buffer recipe

One of the major factor in trouble-shooting streaking is protein solubility (Görg *et al.* 2000). Proteins need to be solubilized effectively throughout the duration of IEF run. An effective solubilizing buffer also increases the number of different proteins analysable in a 2D gel. A variety of 2DE-buffer recipes have been used by researchers over the years. Urea, thiourea, CHAPS and DTT concentrations have already been standardized for *E. coli*; several manipulations have been attempted which did not necessarily result in good 2DE gels (Choe and Lee 2000; Han *et al.* 2005; Herbert *et al.* 2006; Nandakumar *et al.* 2003; Smejkal *et al.* 2006; Valente *et al.* 2012). We suggest a suitable 2DE-buffer recipe, i.e. 7 M urea, 2 M thiourea, 65 mM DTT, 4% CHAPS and 2%

ampholytes. Ampholytes help in solubilization of proteins near their *pI* (Righetti *et al.* 2007; Westermeier and Naven 2002), aids in removing nucleic acids (Rabilloud 1996; Rabilloud *et al.* 1986; Shaw and Riederer 2003) and also scavenges cyanate ions responsible for non-specific modification of proteins (Shaw and Riederer 2003). Generally 0.5% to 2% of ampholyte is recommended (Shaw and Riederer 2003); according to our observations 2% of ampholytes is optimum. We have used molecular biology-grade urea, ordinary-grade thiourea, electrophoresis-grade CHAPS and DTT and dd-Milli-Q. Ultrapure reagents will produce even better quality 2DE gels.

4.2 Removal of NuA contamination and simplified protein extraction

Apart from causing ARS, the large NuA molecules become too viscous to work with, in RB. There are two main strategies generally practiced to remove NuA contamination. One is digestion by nuclease and the other is selective precipitation of proteins from contaminating medium and resolubilization in RB. These are often practiced separately or in tandem in sample preparation of 2DE (Görg 2004; Stasyk *et al.* 2001). The cleanup procedure removes contaminants like salts, NuA and polysaccharides. A strategy involving precipitating proteins in the aqueous-organic interface (Antonioli *et al.* 2009) has been shown to be effective in removing NuA. However, precipitation and resolubilization steps in the cleanup process do not ensure total recovery of proteins. Very often significant amounts of proteins are disproportionately lost during cleanup process especially when multiple rounds of processing is needed, which

Figure 3. Histogram of length of streaking and number of ARS in few recently published literature, including those discussing 2DE optimization with *E. coli* WCPE. The open bars represent cases with at least 10 ARS. The Roman numerals at the x-axis represent the following publications: (i) (Fedyunin *et al.* 2012); (ii) (Zuo and Speicher 2000); (iii) (Valente *et al.* 2012) (Fig. 5E); (iv) (Nandakumar *et al.* 2003); (v) (Cheung *et al.* 2012); (vi) (Choe and Lee 2000); (vii) (Valente *et al.* 2012) (Fig. 5D); (viii) (Qi *et al.* 2013) (Fig. 1B); (ix) (Kim *et al.* 2012) (Fig. 5A); (x) (Kim *et al.* 2012) (Fig. 5B); (xi) (Smejkal *et al.* 2006); (xii) (Herbert *et al.* 2006); (xiii) (Qi *et al.* 2013) (Fig. 1A); (xiv) (Han *et al.* 2005); (xv) (Riley *et al.* 2012) (Fig. 2C); (xvi) (Ramachandran *et al.* 2012); (xvii) (Piras *et al.* 2012); (xviii) (Aich *et al.* 2012); (xix) (Ramachandran *et al.* 2012) (Fig. 5C); (xx) (Schliep *et al.* 2012); (xxi) (Riley *et al.* 2012) (Fig. 2A); (xxii) (Aich *et al.* 2012); (xxiii) This work. Figure numbers were mentioned where more than one 2DE gel image was used.

limits the use of 2DE in quantitative applications. Also, in some cases the acid used for sample precipitation carries over to final 2DE solution producing artifacts. Difficulty in solubilizing the protein precipitate is also a commonly faced problem. Because of the low salt content of *E. coli*, WCPE sample cleanup may be avoided if NuA is eliminated by any other means. A moderate amount of salt contamination may also be addressed by our IEF program and paper wick strategy.

In our study, the cleanup procedure did not achieve efficient removal of NuA, suggesting necessity of multiple rounds of processing. It is likely that we did not get enough nuclease activity in strongly denaturing RB. Besides such an approach adds extra contamination in the form of foreign proteins, salts and other substances (needed for enzyme activity and stability) preventing from producing well focused spots.

We demonstrated that sonication of cells in dd-Milli-Q maintaining ice-cold environment rather than sonicating in RB yields better results. While sonication is a standard technique regularly used in sample preparation, effective sonication is significantly improved by lowering the viscosity of the aqueous medium by using pure water instead of RB, allowing better transmission of ultrasound wave. More rounds of sonication in RB or use of maximum allowed power could not compensate for the advantage associated with low viscosity of water. Presence of detergent in RB causes frothing, which is easily avoided in pure water. Therefore, sonication in less viscous water not only facilitates cell rupture but total shearing of DNA molecules as well. The small amount of low-to-medium sized NuA molecules still present did not pose any serious problem when we optimized other parameters in 2DE protocol. It is pertinent to ask if sonication could collaterally lead to breaking of protein backbones, especially of the large ones. It has been previously shown that such breakage, if at all, is minor and becomes insignificant at higher protein concentrations as it imparts a protective effect against mechanical shear (Coakley *et al.* 1973). This is also corroborated by the supplementary figure 2, lane 2 and 3, depicting gel electrophoresis pattern by our protocol and standard protocol, which show no obvious differences. Furthermore, the sonication step requires at the most 10 min, and considering the benefits of clear 2DE gels anticipated, the protocol offers clear advantage compared to minor increase of work load or time of sample preparation.

The advantages of our protocol are manifold. Foremost of these is the simplified sample preparation protocol involving minimal steps. Simpler procedure means lesser manual error, yielding improved reproducibility both in terms of recoveries and maintenance of stoichiometric amounts of the proteins important for accurate quantitative proteome analyses. It minimizes transfer of proteins from one solution to the other, which helps maintain original protein composition. Our protocol effectively removes IEF-interfering NuA, avoiding any further processing or enzymatic treatment. This eliminates the chance of external contaminants coming into the sample. We

also did not need to expose the proteins for long-time incubation at room temperature, needed for nuclease digestion. Finally, because the procedure is short, it ensures protein stability without the need for addition of protease inhibitors, ensuring minimum modification of the WCPE as well.

4.3 Optimized IEF and paper wick replacement strategy

In recent years, short IEF programs are being supported for their fast processing time. However, in our experience, a 12 h active rehydration and considerably long IEF program gives improved results. Active rehydration with 50 V constant voltage applied between electrodes ensures better absorption of sample in to the IPG strip gel. The IEF program we state (supplementary figure 5) starts with low voltage of 250 to 300 V for few hours, so as to carry the salt and other conducting small molecules into the paper wicks at the ends. As the paper wicks are changed on hourly basis, more and more conducting substances are removed from the IEF system and the current drops. Slowly we increase the voltage step by step and finally, after a considerable salt removal period, we move towards higher voltage (3500 to 8000 V) when protein molecules can be mobilized towards their focal point in the pH gradient on the IPG strip. The program was run till required volt-hours were achieved (dotted lines in supplementary figure 5).

The idea of changing paper wicks was highlighted more than a decade back in 1997 by Görg *et al.* (1997), who observed that changing cathodic paper wick during IEF help to reduce dryness of the gel at the cathodic side of the IPG strip. Lai *et al.* (2003) in 2003 noticed an improvement in focusing when paper wicks were changed frequently together with longer IEF for *Bacillus subtilis* spore sample; but these reports did not provide any detailed information on the wick changing strategy. This may be a possible reason why its use is not as commonly reported in literature despite clear benefits associated to it. In this work we have given detailed criteria for changing paper wicks and showed that regular changing of the paper wicks alongside an optimized IEF program can improve the quality of 2DE gels both in terms of minimal streaking and more spots.

4.4 A robust protocol

M. smegmatis being an acid-fast Gram-positive bacterium has higher fatty acid content and a very different proteome profile and cell composition than *E. coli*. But our OP is quite robust in eliminating ARS when applied to WCPE from diverse organisms. It is noteworthy that a higher amount (twice that of *E. coli*) of initial cell pellet of *M. smegmatis* was required to reach similar final total protein concentrations. It may also be noted that we have demonstrated the use of our protocol in both 7 and 17 cm IPG strips wherein the

IEF program and the paper-wick changing strategy could be scaled using a common principle, suggesting its utility for any length IPG strip.

4.5 Comparative analysis of performance

The comparative analysis showed that our method effectively eliminates ARS from 2DEs of two very different types of bacteria, *E. coli* and *M. smegmatis*. At the same time it improves the number of detected spots in those 2DEs. It may also be noted that there have been publications on 2DE optimization over the years, but there is not much correlation between the year of publication and the extent of streaking. This justifies the high relevance of this work in the field of proteomics.

The two goals of a 2DE protocol are to produce clean 2DE gels without any artifacts, like ARS, and to reveal the protein composition of the sample devoid of any loss or modification (Westermeyer and Naven 2002). We showed that both of the goals can be achieved at the same time with our protocol, demonstrated with *E. coli* and *M. smegmatis* WCPE. Our protocol proposes a suitable 2DE-buffer recipe for WCPE, i.e. 7 M urea, 2 M thiourea, 4% CHAPS, 65 mM DTT and 2% ampholytes. We have greatly simplified the sample preparation procedure by excluding the standard sample cleanup, while maintaining protein recovery and avoiding unnecessary contaminants. We present an optimized IEF program for well-resolved IEF of proteins eliminating ARS. Our methodology will simplify the procedure, thereby allowing cost-effective, time-efficient and improved-quality 2DE gels.

Acknowledgements

This work was supported by the Department of Biotechnology, New Delhi, India.

References

- Aich P, Patra M, Chatterjee AK, Roy SS and Basu T 2012 Calcium chloride made *E. coli* competent for uptake of extraneous DNA through overproduction of OmpC protein. *Protein J.* **31** 366–373
- Amersham-Biosciences 2003 2-D Clean-pp kit (Amersham Biosciences)
- Antonoli P, Bachi A, Fasoli E and Righetti PG 2009 Efficient removal of DNA from proteomic samples prior to two-dimensional map analysis. *J. Chromatogr. A* **1216** 3606–3612
- Bai F, Liu S and Witzmann FA 2005 A 'de-streaking' method for two-dimensional electrophoresis using the reducing agent tris(2-carboxyethyl)-phosphine hydrochloride and alkylating agent vinylpyridine. *Proteomics* **5** 2043–2047
- Bio-Rad 2011 ReadyStrip IPG Strip. Instruction Manual
- Bradford MM 1976 A rapid and sensitive method for the quantitation of microgram quantities of protein utilizing the principle of protein-dye binding. *Anal. Biochem.* **72** 248–254
- Butt AQ, Ahmed S, Maratha A and Miggin SM 2012 14-3-3epsilon and 14-3-3sigma inhibit Toll-like receptor (TLR)-mediated proinflammatory cytokine induction. *J. Biol. Chem.* **287** 38665–38679
- Chen YZ, Pang QY, He Y, Zhu N, Branstrom I, Yan XF and Chen S 2012 Proteomics and metabolomics of Arabidopsis responses to perturbation of glucosinolate biosynthesis. *Mol. Plant* **5** 1138–1150
- Cheung HY, Wong MM, Cheung SH, Liang LY, Lam YW and Chiu SK 2012 Differential actions of chlorhexidine on the cell wall of *Bacillus subtilis* and *Escherichia coli*. *PLoS One* **7** e36659
- Chiangjong W and Thongboonkerd V 2009 A comparative study of different dyes for the detection of proteomes derived from *Escherichia coli* and MDCK cells: sensitivity and selectivity. *J. Chromatogr. B. Anal. Technol. Biomed. Life Sci.* **877** 1433–1439
- Choe LH and Lee KH 2000 A comparison of three commercially available isoelectric focusing units for proteome analysis: the multiphor, the IPGphor and the protean IEF cell. *Electrophoresis* **21** 993–1000
- Coakley WT, Brown RC, James CJ and Gould RK 1973 The inactivation of enzymes by ultrasonic cavitation at 20 kHz. *Archives Biochem. Biophys.* **159** 722–729
- Cremer F and Van de Walle C 1985 Method for extraction of proteins from green plant tissues for two-dimensional polyacrylamide gel electrophoresis. *Anal. Biochem.* **147** 22–26
- Damerval C, De Vienne D, Zivy M and Thiellement H 1986 Technical improvements in two-dimensional electrophoresis increase the level of genetic variation detected in wheat-seedling proteins. *Electrophoresis* **7** 52–54
- Fedyunin I, Lehnhardt L, Bohmer N, Kaufmann P, Zhang G and Ignatova Z 2012 tRNA concentration fine tunes protein solubility. *FEBS Lett* **586** 3336–3340
- Görg A 2004 2-D Electrophoresis principles and methods. *GE Healthcare* 25–29
- Görg A, Obermaier C, Boguth G, Csordas A, Diaz JJ and Madjar JJ 1997 Very alkaline immobilized pH gradients for two-dimensional electrophoresis of ribosomal and nuclear proteins. *Electrophoresis* **18** 328–337
- Görg A, Obermaier C, Boguth G, Harder A, Scheibe B, Wildgruber R and Weiss W 2000 The current state of two-dimensional electrophoresis with immobilized pH gradients. *Electrophoresis* **21** 1037–1053
- Görg A, Postel W, Weser J, Günther S, Strahler JR, Hanash SM and Somerlot L 1987 Elimination of point streaking on silver stained two-dimensional gels by addition of iodoacetamide to the equilibration buffer. *Electrophoresis* **8** 122–124
- Han MJ, Lee JW and Lee SY 2005 Enhanced proteome profiling by inhibiting proteolysis with small heat shock proteins. *J. Proteome Res.* **4** 2429–2434
- Hatairaktham S, Srisawat C, Siritanaratkul N, Chiangjong W, Fucharoen S, Thongboonkerd V and Kalpravidh RW 2013 Differential plasma proteome profiles of mild versus severe beta-thalassemia/Hb E. *Ann. Hematol.* **92** 365–377
- Havanapan PO and Thongboonkerd V 2009 Are protease inhibitors required for gel-based proteomics of kidney and urine? *J. Proteome Res.* **8** 3109–3117

- Henzel WJ, Billeci TM, Stults JT, Wong SC, Grimley C and Watanabe C 1993 Identifying proteins from two-dimensional gels by molecular mass searching of peptide fragments in protein sequence databases. *Proc. Natl. Acad. Sci. USA* **90** 5011–5015
- Herbert BR, Grinyer J, McCarthy JT, Isaacs M, Harry EJ, Nevalainen H, Traini MD, Hunt S *et al.* 2006 Improved 2-DE of microorganisms after acidic extraction. *Electrophoresis* **27** 1630–1640
- Hong Y, Sun A, Zhang M, Gao F, Han Y, Fu Z, Shi Y and Lin J 2013 Proteomics analysis of differentially expressed proteins in schistosomula and adult worms of *Schistosoma japonicum*. *Acta Trop.* **126** 1–10
- James P, Quadroni M, Carafoli E and Gonnet G 1993 Protein identification by mass profile fingerprinting. *Biochem. Biophys. Res. Commun.* **195** 58–64
- Jiang T, Zhan Y, Sun M, Liu S, Zang S, Ma Y and Xin Y 2011 The novel responses of ethambutol against *Mycobacterium smegmatis* mc(2)155 revealed by proteomics analysis. *Curr. Microbiol.* **62** 341–345
- Kamthan M, Mukhopadhyay G, Chakraborty N, Chakraborty S and Datta A 2012 Quantitative proteomics and metabolomics approaches to demonstrate N-acetyl-D-glucosamine inducible amino acid deprivation response as morphological switch in *Candida albicans*. *Fungal Genet. Biol.* **49** 369–378
- Kao S-H, Wong H-K, Chiang C-Y and Chen H-M 2008 Evaluating the compatibility of three colorimetric protein assays for two-dimensional electrophoresis experiments. *Proteomics* **8** 2178–2184
- Khoudoli G, Porter I, Blow J and Swedlow J 2004 Optimisation of the two-dimensional gel electrophoresis protocol using the Taguchi approach. *Proteome Sci.* **2** 6
- Kim Y, Lee JW, Kang SG, Oh S and Griffiths MW 2012 Bifidobacterium spp. influences the production of autoinducer-2 and biofilm formation by *Escherichia coli* O157:H7. *Anaerobe* **18** 539–545
- Lai EM, Phadke ND, Kachman MT, Giorno R, Vazquez S, Vazquez JA, Maddock JR and Driks A 2003 Proteomic analysis of the spore coats of *Bacillus subtilis* and *Bacillus anthracis*. *J. Bacteriol.* **185** 1443–1454
- Liao H, Zhang F, Hu X and Liao X 2011 Effects of high-pressure carbon dioxide on proteins and DNA in *Escherichia coli*. *Microbiol.* **157** 709–720
- Lopez JL 2007 Two-dimensional electrophoresis in proteome expression analysis. *J. Chromatogr. B Analyt. Technol. Biomed. Life Sci.* **849** 190–202
- Low B 1968 Formation of merodiploids in matings with a class of Rec⁻ recipient strains of *Escherichia coli* K12. *Proc. Natl. Acad. Sci. USA* **60** 160–167
- Nakata K, Koh MM, Tsuchido T and Matsumura Y 2010 All genomic mutations in the antimicrobial surfactant-resistant mutant, *Escherichia coli* OW66, are involved in cell resistance to surfactant. *Appl. Microbiol. Biotechnol.* **87** 1895–1905
- Nandakumar MP, Shen J, Raman B and Marten MR 2003 Solubilization of trichloroacetic acid (TCA) precipitated microbial proteins via NaOH for two-dimensional electrophoresis. *J. Proteome Res.* **2** 89–93
- O'Farrell PH 1975 High resolution two-dimensional electrophoresis of proteins. *J. Biol. Chem.* **250** 4007–4021
- Olsson I, Larsson K, Palmgren R and Bjellqvist B 2002 Organic disulfides as a means to generate streak-free two-dimensional maps with narrow range basic immobilized pH gradient strips as first dimension. *Proteomics* **2** 1630–1632
- Pal P, Kanaujiya JK, Lochab S, Tripathi SB, Sanyal S, Behre G and Trivedi AK 2013 Proteomic analysis of rosiglitazone and guggulsterone treated 3T3-L1 preadipocytes. *Mol. Cell Biochem.* **376** 81–93
- Pappin DJ, Hojrup P and Bleasby AJ 1993 Rapid identification of proteins by peptide-mass fingerprinting. *Curr. Biol.* **3** 327–332
- Pennington K, McGregor E, Beasley CL, Everall I, Cotter D and Dunn MJ 2004 Optimization of the first dimension for separation by two-dimensional gel electrophoresis of basic proteins from human brain tissue. *Proteomics* **4** 27–30
- Piras C, Soggiu A, Bonizzi L, Gaviraghi A, Deriu F, De Martino L, Iovane G, Amoresano A 2012 Comparative proteomics to evaluate multi drug resistance in *Escherichia coli*. *Mol. Biosyst.* **8** 1060–1067
- Qi J, Du Y, Bai H, Zhu X, Hu M, Luo Y and Liu Y 2013 Global protein expression profile response of *Escherichia coli* ATCC 25922 exposed to enrofloxacin. *Microb. Drug Resist.* **19** 6–14
- Rabilloud T 1996 Solubilization of proteins for electrophoretic analyses. *Electrophoresis* **17** 813–829
- Rabilloud T 2002 Two-dimensional gel electrophoresis in proteomics: old, old fashioned, but it still climbs up the mountains. *Proteomics* **2** 3–10
- Rabilloud T, Chevallet M, Luche S and Lelong C 2010 Two-dimensional gel electrophoresis in proteomics: Past, present and future. *J. Proteomics* **73** 2064–2077
- Rabilloud T, Hubert M and Tarroux P 1986 Procedures for two-dimensional electrophoretic analysis of nuclear proteins. *J. Chromatogr. A* **351** 77–89
- Rajani DK, Walch M, Martinvalet D, Thomas MP and Lieberman J 2012 Alterations in RNA processing during immune-mediated programmed cell death. *Proc. Natl. Acad. Sci. USA* **109** 8688–8693
- Ramachandran B, Dikshit KL and Dharmalingam K 2012 Recombinant *E. coli* expressing Vitreoscilla haemoglobin prefers aerobic metabolism under microaerobic conditions: a proteome-level study. *J. Biosci.* **37** 617–633
- Righetti PG, Simo C, Sebastiano R and Citterio A 2007 Carrier ampholytes for IEF, on their fortieth anniversary (1967–2007), brought to trial in court: the verdict. *Electrophoresis* **28** 3799–3810
- Riley LM, Veses-Garcia M, Hillman JD, Handfield M, McCarthy AJ and Allison HE 2012 Identification of genes expressed in cultures of *E. coli* lysogens carrying the Shiga toxin-encoding prophage Phi24B. *BMC Microbiol.* **12** 42
- Schliep M, Ryall B and Ferenci T 2012 The identification of global patterns and unique signatures of proteins across 14 environments using outer membrane proteomics of bacteria. *Mol. Biosyst.* **8** 3017–3027
- Shaw MM and Riederer BM 2003 Sample preparation for two-dimensional gel electrophoresis. *Proteomics* **3** 1408–1417
- Shires K and Steyn L 2001 The cold-shock stress response in *Mycobacterium smegmatis* induces the expression of a histone-like protein. *Mol. Microbiol.* **39** 994–1009

- Smejkal GB, Robinson MH, Lawrence NP, Tao F, Saravis CA and Schumacher RT 2006 Increased protein yields from *Escherichia coli* using pressure-cycling technology. *J. Biomol. Tech.* **17** 173–175
- Stasyk T, Hellman U and Souhelnytskyi S 2001 Optimizing sample preparation for 2-D electrophoresis. *Life Sci. News* **9** 8–11
- Sun H, Zhang A, Yan G, Han Y, Sun W, Ye Y and Wang X 2013 Proteomics study on the hepatoprotective effects of traditional Chinese medicine formulae Yin-Chen-Hao-Tang by a combination of two-dimensional polyacrylamide gel electrophoresis and matrix-assisted laser desorption/ionization-time of flight mass spectrometry. *J. Pharm. Biomed. Anal.* **75** 173–179
- Sun J, Li H and Yuan Q 2012 Metabolic regulation of trisporic acid on *Blakeslea trispora* revealed by a GC-MS-based metabolomic approach. *PLoS One* **7** e46110
- Thiede B, Koehler CJ, Strozynski M, Treumann A, Stein R, Zimny-Arndt U, Schmid M and Jungblut PR 2013 High resolution quantitative proteomics of HeLa cells protein species using stable isotope labeling with amino acids in cell culture (SILAC), two-dimensional gel electrophoresis (2DE) and nano-liquid chromatography coupled to an LTQ-OrbitrapMass spectrometer. *Mol. Cell Proteomics* **12** 529–538
- Valente KN, Choe L, Lenhoff A and Lee K 2012 Optimization of protein sample preparation for two-dimensional electrophoresis. *Electrophoresis* **33** 1947–1957
- Westermeyer R and Naven T 2002 *Proteomics in practice* (Weinheim: Wiley-VCH) p 19
- Xiao H, Zhang L, Zhou H, Lee JM, Garon EB and Wong DT 2012 Proteomic analysis of human saliva from lung cancer patients using two-dimensional difference gel electrophoresis and mass spectrometry. *Mol. Cell Proteomics* **11** M111 012112
- Yates JR 3rd, Speicher S, Griffin PR and Hunkapiller T 1993 Peptide mass maps: a highly informative approach to protein identification. *Anal. Biochem.* **214** 397–408
- Yun H, Lee JW, Jeong J, Chung J, Park JM, Myoung HN and Lee SY 2007 EcoProDB: the *Escherichia coli* protein database. *Bioinformatics* **23** 2501–2503
- Zhang B, Xu C, Zhang L, Zhou S, Feng S, He Y and Liao M 2013 Enhanced adherence to and invasion of PUVEC and PK-15 cells due to the overexpression of RfaD, ThyA and Mip in the DeltaompP2 mutant of *Haemophilus parasuis* SC096 strain. *Vet. Microbiol.* **162** 713–723
- Zhang Q, Lu YX and Xu WH 2012 Integrated proteomic and metabolomic analysis of larval brain associated with diapause induction and preparation in the cotton bollworm, *Helicoverpa armigera*. *J. Proteome Res.* **11** 1042–1053
- Zuo X and Speicher DW 2000 Quantitative evaluation of protein recoveries in two-dimensional electrophoresis with immobilized pH gradients. *Electrophoresis* **21** 3035–3047

MS received 14 February 2014; accepted 31 May 2014

Corresponding editor: B JAGADEESHWAR RAO